

DE 3 - 31905

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

**CB TEST CERTIFICATE** 

Product Laboratory Equipment

(microscope)

Name and address of the applicant Leica Microsystems CMS GmbH

Ernst-Leitz-Straße 17-37

35578 Wetzlar GERMANY

Name and address of the manufacturer

Leica Microsystems CMS GmbH

Ernst-Leitz-Straße 17-37, 35578 Wetzlar, GERMANY

Name and address of the factory

Leica Microsystems CMS GmbH

Ernst-Leitz-Straße 17-37, 35578 Wetzlar, GERMANY

Ratings and principal characteristics Rated voltage: 100-240 V Rated frequency: 50/60 Hz

Rated frequency: 50/60 H.
Rated power: 14 VA
Protection class: I

Conditions of acceptability:

- This equipment is for indoor use in non-hazardous locations, operated by qualified

personnel.

- The detachable power cord set shall comply with the National Standards and/or National Regulations of the country in question and/or in compliance with this

standard.

- The detachable power cord set length shall be not more than 3  $\mbox{m}.$ 

Trademark / Brand (if any) Leica

Model/type Ref. DMIL LED, DMIL LED Fluo, DMIL LED Cellfactory,

DMIL Fluo LED Cellfactory, DMIL M LED

Additional information (if necessary)

Certificate DE 3 – 31710 issued 2022-02-07 is replaced by this version due

to technical changes

A sample of the product was tested and found

to be in conformity with

IEC 61010-1:2010

IEC 61010-1:2010/AMD1:2016

IEC 61010-2-101:2018

as shown in the Test Report Ref. No. which forms part of this certificate

713275729

Page 1 of 1

This CB Test Certificate is issued by the National Certification Body

CB 056257 0283 Rev. 01

Date, 2022-11-15

Es Ollan

Product Service