







CERTIFICATE

No. U8 056257 0284 Rev. 01

Holder of Certificate: Leica Microsystems CMS GmbH

Ernst-Leitz-Straße 17-37

35578 Wetzlar GERMANY

Certification Mark:



Product: Laboratory Equipment

(microscope)

Model(s): DMIL LED, DMIL LED Fluo, DMIL LED Cellfactory,

DMIL Fluo LED Cellfactory, DMIL M LED

Parameters: Rated voltage: 100-240 V

Rated frequency: 50/60 Hz Rated power: 14 VA Protection class: I

Conditions of acceptability:

- This equipment is for indoor use in non-hazardous locations, operated by qualified personnel.

 The detachable power cord set shall comply with the National Standards and/or National Regulations of the country in question and/or in compliance with this standard.

- The detachable power cord set length shall be not more than 3 m.

TestedUL 61010-1:2012/R:2019-07

according to: UL 61010-2-101:2019

CSA C22.2 No. 61010-1:2012/A1:2018-11 CSA C22.2 No. 61010-2-101:2019

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL for USA and a Standards Council of Canada ISO/IEC 17065 accredited Certification body for Canada.

Test report no.: 713275729

Date, 2022-11-15

(Erich Thurner)