



CERTIFICATE

No. Z1 15 09 56257 134

Holder of Certificate: Leica Microsystems CMS GmbH

Ernst-Leitz-Straße 17-37

35578 Wetzlar **GERMANY**

Production

Product:

56257

Facility(ies):

Certification Mark:



Laboratory Equipment

(Microscope)

Model(s): DMIL M LED

DMIL LED Fluo

(Microscope for In-Vitro-Diagnostic (IVD))

DMIL LED

(Microscope for In-Vitro-Diagnostic (IVD))

DMIL LED Cellfactory

(Microscope for In-Vitro-Diagnostic (IVD))

DMIL LED Fluo Cellfactory

(Microscope for In-Vitro-Diagnostic (IVD))

Parameters: Rated Voltage: 100-240V

Rated Frequency: 50/60Hz Rated Input Power: 14 VA Protection Class:

Tested according to: EN 61010-1:2010

EN 61010-1:2001 EN 62471:2008 EN 61010-2-101:2002

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

Test report no.: 028-713067113-000

Valid until: 2020-09-06

(Ralph Fischer)

Page 1 of 1

2015-09-08

Date,





