



America

CERTIFICATE

No. U8 056257 0278 Rev. 00

Holder of Certificate: **Leica Microsystems CMS GmbH**
Ernst-Leitz-Straße 17-37
35578 Wetzlar
GERMANY

Certification Mark:



Product: **Laboratory Equipment
(Microscope)**

Model(s): **DM1000 LED**

Parameters:
Rated voltage: 12 VDC
Rated input power: 18 W
Protection class: III

Tested according to:
UL 61010-1:2012/R:2019-07
UL 61010-2-101:2019
CSA C22.2 No. 61010-1:2012/A1:2018-11
CSA C22.2 No. 61010-2-101:2019

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited Certification body.

Test report no.: 713229324

Date, 2022-01-28

(Erich Thurner)