

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT  
(IECEE) CB SCHEME

## CB TEST CERTIFICATE

|  |   |
|--|---|
| Product  | Laboratory Equipment<br>(microscope)  |
| Name and address of the applicant  | <b>Leica Microsystems CMS GmbH</b><br>Ernst-Leitz-Straße 17-37<br>35578 Wetzlar<br>GERMANY  |
| Name and address of the manufacturer   | Leica Microsystems CMS GmbH<br>Ernst-Leitz-Straße 17-37, 35578 Wetzlar, GERMANY   |
| Name and address of the factory  | Leica Microsystems Ltd. Shanghai<br>Floor 1, 2, 3A, 4A, and 6, Building T20-1, 258 Jinzang Road, China<br>(Shanghai) Pilot Free Trade Zone, 201206 Shanghai, PEOPLE'S<br>REPUBLIC OF CHINA<br><br>Zhang Jia Gang Optical Instrument Co; Ltd<br>Tangqiao Town, Luyuan Nanlin Development Zone, 215616 Zhangjiagang<br>City, PEOPLE'S REPUBLIC OF CHINA |
| Ratings and principal characteristics  | Rated voltage: 100-240 V<br>Rated frequency: 50/60 Hz<br>Rated power: 75 VA<br>Protection class: I  |
| Trade mark (if any)  | Leica   |
| Model/type Ref.  | DM1000  |
| Additional information (if necessary)  | Certificate DE 3 – 31141 issued 2020-01-14 is replaced by this version due<br>to technical changes  |
| A sample of the product was tested and found<br>to be in conformity with     | IEC 61010-1:2010<br>IEC 61010-1:2010/AMD1:2016<br>IEC 61010-2-101:2018  |
| as shown in the Test Report Ref. No.<br>which forms part of this certificate | 713228732   |

Page 1 of 1

This CB Test Certificate is issued by the National Certification Body

CB 056257 0273 Rev. 00

Date, 2022-01-26



( Ralph Fischer )

TÜV SÜD Product Service GmbH • Certification Body • Ridlerstraße 65 • 80339 Munich • Germany



Product Service