



Ref. Certif. No.

DE 3 - 31920

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT  
(IECEE) CB SCHEME

## CB TEST CERTIFICATE

Product

Laboratory Equipment  
(Microscope)

Name and address of the applicant

**Leica Microsystems CMS GmbH**  
Ernst-Leitz-Straße 17-37  
35578 Wetzlar  
GERMANY

Name and address of the manufacturer

Leica Microsystems CMS GmbH  
Ernst-Leitz-Straße 17-37, 35578 Wetzlar, GERMANY

Name and address of the factory

Leica Microsystems CMS GmbH  
Ernst-Leitz-Straße 17-37, 35578 Wetzlar, GERMANY

Ratings and principal characteristics

Rated voltage:	100-240 V
Rated frequency:	50/60 Hz
Rated power:	100 VA
Protection class:	I
Ambient temp.:	restricted, 15 to 35 °C, max. 80 % r.H. @ 30 °C
Pollution degree:	2
Overvoltage category:	II

Trademark / Brand (if any)

Leica

Model/type Ref.

**DM 8000 M, DM 12000 M**A sample of the product was tested and found  
to be in conformity withIEC 61010-1:2010  
IEC 61010-1:2010/AMD1:2016as shown in the Test Report Ref. No.  
which forms part of this certificate

713259670

Page 1 of 1

This CB Test Certificate is issued by the National Certification Body

CB 056257 0305 Rev. 00

Date, 2022-12-09

( Ralph Fischer )

TÜV SÜD Product Service GmbH • Certification Body • Ridlerstraße 65 • 80339 Munich • Germany



Product Service