CERTIFICATE

No. Z1 18 05 56257 179

Holder of Certificate: Leica Microsystems CMS GmbH

Ernst-Leitz-Straße 17-37

35578 Wetzlar **GERMANY**

Production Facility(ies):

30821

Certification Mark:



Laboratory Equipment Product:

(microscope for IVD)

Model(s): **DM2000 LED**

Type: 11888843

Parameters: 100-240 V Rated voltage:

50/60 Hz Rated frequency: 30 VA Rated power: Protection class:

Tested according to: EN 61010-1:2010

EN 61010-2-101:2017

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

Test report no.: 028-713133754-000

Valid until: 2023-05-16

Date, 2018-05-22 (Ralph Fischer)

Page 1 of 1



