



Product Service

# C E R T I F I C A T E

No. Z1 15 05 56257 115

**Holder of Certificate:** Leica Microsystems CMS GmbHErnst-Leitz-Straße 17-37  
35578 Wetzlar  
GERMANY**Production  
Facility(ies):**

56257

**Certification Mark:****Product:****Laboratory equipment  
(Microscope)****Model(s):****DM6 M, DM6 B, DM6 Q, DM6 CS****Parameters:**Rated voltage: 12 VDC  
Protection class: III

Supplied via control unit CTR6 LED or control unit CTR6

DM6 B: 11888857, 11888859, 11888872, 11888860,  
11888865, 11888864, 11888863, 11888862

DM6 M: 11888777, 11888785, 11888784, 11888776

DM6 Q: 11888880

DM6 CS: 11888413

**Tested according to:**EN 61010-1:2010  
EN 61010-1:2001  
EN 61010-2-101:2002

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

**Test report no.:**

028-713059251-000

**Valid until:**

2020-05-19

*i.A. Ralph Fischer***Date,** 2015-05-22

(Ralph Fischer)

Page 1 of 1

