



America

CERTIFICATE

No. U8 15 06 56257 125

Holder of Certificate: Leica Microsystems CMS GmbH

Ernst-Leitz-Straße 17-37
35578 Wetzlar
GERMANY

**Production
Facility(ies):**

56257

Certification Mark:



C US

Product:

Laboratory Equipment

Model(s):

DM6 FS, DM6 CFS

Parameters:

Model reference:

DM6 FS

DM6 CFS

Type reference:

11888867

11888881

Supplied via control unit CTR6

Rated voltage:

12 VDC

Protection class:

III

**Tested
according to:**

UL 61010-1:2012-05

CAN/CSA-C22.2 no. 61010-1:2012-05

The product was voluntarily tested according to the relevant safety requirements noted above. It can be marked with the certification mark above. The mark must not be altered in anyway. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC Guide 67. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited certification body.

Test report no.:

028-713059248-000

Date, 2015-06-24

Page 1 of 1

