



Product Service

CERTIFICATE

No. Z1 15 06 56257 124

Holder of Certificate: Leica Microsystems CMS GmbH

 Ernst-Leitz-Straße 17-37
 35578 Wetzlar
 GERMANY

**Production
Facility(ies):**

56257

Certification Mark:

Product: Laboratory Equipment

Model(s): DM6 FS, DM6 CFS

Parameters:

 Model reference:
 DM6 FS
 DM6 CFS

 Type reference:
 11888867
 11888881

Supplied via control unit CTR6

 Rated voltage: 12 VDC
 Protection class: III

Tested according to: EN 61010-1:2010
 EN 62471:2008

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

Test report no.: 028-713059248-000

Valid until: 2020-06-22

Date, 2015-06-24

(Ralph Fischer)

Page 1 of 1

