



IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

**CB TEST CERTIFICATE** 

Product Laboratory Equipment

(microscope)

Name and address of the applicant Leica Microsystems CMS GmbH

Ernst-Leitz-Straße 17-37

35578 Wetzlar GERMANY

Name and address of the manufacturer Leica Microsystems CMS GmbH

Ernst-Leitz-Straße 17-37, 35578 Wetzlar, GERMANY

Name and address of the factory

Leica Microsystems CMS GmbH

Ernst-Leitz-Straße 17-37, 35578 Wetzlar, GERMANY

Ratings and principal characteristics Microscope:

Equipment is supplied via control unit CTR Compact, CTR Compact XY, CTR Advanced or CTR Advanced+.

Leica LIBS Module: 24 Vdc, 50 W, class III

Trade mark (if any) Leica

Model/type Ref. Model: DM6 B

Type: 11889186, 11889187

Model: DM6 M

Type: 11889191, 11889192, 11889193, 11889194

Model: DM6 CS Type: 11889199

Model: Leica LIBS Module

Type: 11560200

A sample of the product was tested and found

to be in conformity with

IEC 61010-1:2010

IEC 61010-1:2010/AMD1:2016

IEC 60825-1:2014

as shown in the Test Report Ref. No. which forms part of this certificate

713212608-02

Page 1 of 1

This CB Test Certificate is issued by the National Certification Body

CB 056257 0243 Rev. 00

Date, 2021-07-26

(Benedikt Pulver)

TÜV SÜD Product Service GmbH • Certification Body • Ridlerstraße 65 • 80339 Munich • Germany



Product Service