

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Control units
Name and address of the applicant	Leica Microsystems CMS GmbH Ernst-Leitz-Straße 17-37 35578 Wetzlar GERMANY
Name and address of the manufacturer	Leica Microsystems CMS GmbH Ernst-Leitz-Straße 17-37, 35578 Wetzlar, GERMANY
Name and address of the factory	Leica Microsystems CMS GmbH Ernst-Leitz-Straße 17-37, 35578 Wetzlar, GERMANY
Ratings and principal characteristics	CTR6 LED: 100-240 V, 50/60 Hz, 150 VA, class I CTR6: 100-240 V, 50/60 Hz, 290 VA, class I
Trade mark (if any)	Leica
Model/type Ref.	CTR6 LED, CTR6
Additional information (if necessary)	Certificate DE 3 - 31016 issued 2019-08-14 is replaced by this version due to technical changes.
A sample of the product was tested and found to be in conformity with	IEC 61010-1:2010 IEC 61010-1:2010/AMD1:2016
as shown in the Test Report Ref. No. which forms part of this certificate	713229356

Page 1 of 1

This CB Test Certificate is issued by the National Certification Body

CB 056257 0281 Rev. 00

Date, 2022-02-03



(Thorsten Siemon)

TÜV SÜD Product Service GmbH • Certification Body • Ridlerstraße 65 • 80339 Munich • Germany



Product Service