



America

# CERTIFICATE

No. U8 056257 0282 Rev. 00

**Holder of Certificate:** **Leica Microsystems CMS GmbH**

Ernst-Leitz-Straße 17-37  
35578 Wetzlar  
GERMANY

**Certification Mark:**



**Product:** **Control units**

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited Certification body.

**Test report no.:** 713229356

**Date,** 2022-02-03

Siemon

( Thorsten Siemon )



America

# CERTIFICATE

No. U8 056257 0282 Rev. 00

**Model(s):** CTR6 LED, CTR6

**Brand Name:** Leica

**Tested according to:** UL 61010-1:2012/R:2019-07  
CSA C22.2 No. 61010-1:2012/A1:2018-11

## Parameters:

**CTR6 LED:**  
100-240 V, 50/60 Hz, 150 VA, class I

**CTR6:**  
100-240 V, 50/60 Hz, 290 VA, class I

## Conditions of acceptability:

- This equipment is for indoor use in non-hazardous locations, operated by qualified personnel skilled in its use.
- The detachable power cord set shall comply with the National Standards and/or National Regulations of the country in question and/or in compliance with this standard.
- The detachable power supply cord length of the equipment shall not be more than 3 m.