



## CERTIFICATE

No. U8 056257 0194 Rev. 00

Holder of Certificate: Leica Microsystems CMS GmbH

030821

Ernst-Leitz-Straße 17-37

35578 Wetzlar GERMANY

**Production** 

Facility(ies):

**Certification Mark:** 



Product: Microscopes

(IVD equipment)

Model(s): DMi1

Parameters:

Rated Voltage: 100-240 VAC
Rated frequency: 50/60 Hz
Rated nower: 18 VA

Rated power: 18 VA
Protection class:

Lamp classification group: exempt group

**Tested** CAN/CSA-C22.2 No. 61010-1:2012/U2:2016-04

according to: CAN/CSA-C22.2 No. 61010-2-101:2015

UL 61010-2-101:2015 UL 61010-1:2012/R:2016-04

The product was voluntarily tested according to the relevant safety requirements noted above. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited certification body.

**Test report no.:** 028-713137139-000

Date, 2018-07-19

Ralph Fischer)