



## CERTIFICATE

No. Z1 056257 0192 Rev. 00

**Holder of Certificate:** Leica Microsystems CMS GmbH

Ernst-Leitz-Straße 17-37

35578 Wetzlar **GERMANY** 030821

**Production** 

Facility(ies):

**Certification Mark:** 



**Product: Microscopes** 

(IVD equipment)

DMi1 Model(s):

Rated Voltage: 100-240 VAC Parameters: Rated frequency: 50/60 Hz

Rated power: 18 VA Protection class:

Lamp classification group: exempt group

EN 61010-1:2010 Tested according to: EN 61010-2-101:2017

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

Test report no.: 028-713137139-000

Valid until: 2023-07-15

Date, 2018-07-19 (Ralph Fischer)