



CERTIFICATE

No. U8 056257 0211 Rev. 00

Holder of Certificate: Leica Microsystems CMS GmbH

Ernst-Leitz-Straße 17-37

35578 Wetzlar **GERMANY**

Production Facility(ies): 056257

Certification Mark:



Product: Laboratory Equipment

(microscope)

DMi8 coded, DMi8 manual, DMi8 automated, Model(s):

DMi8 M, DMi8 C, DMi8 A

CTR compact, CTR advanced or CTR advanced+

Model: DMi8 coded Parameters: DMi8 manual DMi8 M DMi8 C Type: 11889111 11889110 11889092 11889117

100-240 V, 50/60 Hz, 55 VA, Class I

Model: DMi8 automated DMi8 A

Type: 11889112, 11889113, 11889014 11889121, 11889122 Supplied via power supply box CTR compact, CTR advanced or CTR advanced+

UL 61010-1:2012/R2:2016-04 **Tested**

CAN/CSA-C22.2 No. 61010-1:2012/U2:2016-04 according to:

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited Certification body.

Test report no.: 028-713158087-000

Date, 2019-05-02

(Ralph Fischer)