



Ref. Certif. No.

DE 3 - 32537

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Laboratory Equipment (Microscopes)
Name and address of the applicant	Leica Microsystems (Schweiz) AG Max Schmidheiny-Str. 201 9435 Heerbrugg SWITZERLAND
Name and address of the manufacturer	Leica Microsystems (Schweiz) AG Max Schmidheiny-Str. 201, 9435 Heerbrugg, SWITZERLAND
Name and address of the factory	Leica Instruments (Singapore) Pte. Ltd. 15 Tukang Innovation Drive, Singapore 618299, SINGAPORE
	Zhangjiagang Optical Instrument Co., Ltd Nanlin Development Zone, 215616 Zhangjiagang City, PEOPLE'S REPUBLIC OF CHINA
	Shanghai Winsun Power Supply Technology Co. Ltd. Minhang District, No. 110, Wantai Road, Pujiang Town, 201112 Shanghai, PEOPLE'S REPUBLIC OF CHINA
Ratings and principal characteristics	Voltage: 100-240 VAC Frequency: 50 / 60 Hz Wattage: max. 15 W for EZ4, max. 30 W for EZ4 E, EZ4 W
Trademark / Brand (if any)	Leica
Model/type Ref.	EZ4 EZ4 E EZ4 W
Additional information (if necessary)	Certificate DE 3 - 32409 issued 2024-11-15 is replaced by this version due to technical changes. Evaluated to National and Group Differences for US, CA, EU
A sample of the product was tested and found to be in conformity with	IEC 61010-1:2010 IEC 61010-1:2010/AMD1:2016
as shown in the Test Report Ref. No. which forms part of this certificate	713356763

Page 1 of 1

This CB Test Certificate is issued by the National Certification Body

CB 042849 0009 Rev. 03
Date, 2025-04-28

Siemon

(Thorsten Siemon)

TÜV SÜD Product Service GmbH • Certification Body • Ridlerstraße 65 • 80339 Munich • Germany



Product Service