

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT  
(IECEE) CB SCHEME

## CB TEST CERTIFICATE


Product	Laboratory Equipment (comparison macroscope)
Name and address of the applicant	<b>Leica Microsystems CMS GmbH</b> Ernst-Leitz-Straße 17-37 35578 Wetzlar GERMANY
Name and address of the manufacturer	Leica Microsystems CMS GmbH Ernst-Leitz-Straße 17-37, 35578 Wetzlar, GERMANY
Name and address of the factory	Leica Microsystems CMS GmbH Ernst-Leitz-Straße 17-37, 35578 Wetzlar, GERMANY
Ratings and principal characteristics	FS C: 100-240 V, 50/60 Hz, 125 VA (macroscope stand) 12 VDC, 18 VA (comparison bridge, supplied by microscope stand) FS CB: 12 VDC, 18 VA (supplied with external power supply unit)
Trade mark (if any)	Leica
Model/type Ref.	FS C (Macroscope stand with table and comparison bridge) FS CB (comparison bridge)
Additional information (if necessary)	Certificate DE 3 – 3920 issued 2014-11-12 is replaced by this version due to technical changes
A sample of the product was tested and found to be in conformity with	IEC 61010-1:2010 IEC 61010-1:2010/AMD1:2016
as shown in the Test Report Ref. No. which forms part of this certificate	713228734

Page 1 of 1

This CB Test Certificate is issued by the National Certification Body

CB 056257 0275 Rev. 00

Date, 2022-01-26



( Ralph Fischer )

TÜV SÜD Product Service GmbH • Certification Body • Ridlerstraße 65 • 80339 Munich • Germany



Product Service