




America

# CERTIFICATE

No. U8 056257 0276 Rev. 00

**Holder of Certificate:** **Leica Microsystems CMS GmbH**  
Ernst-Leitz-Straße 17-37  
35578 Wetzlar  
GERMANY

**Certification Mark:**



**Product:** **Laboratory Equipment  
(comparison microscope)**

**Model(s):** **FS C (Macroscopic stand with table and comparison bridge)  
FS CB (comparison bridge)**

**Parameters:** FS C: 100-240 V, 50/60 Hz, 125 VA (macroscopic stand)  
12 VDC, 18 VA (comparison bridge, supplied  
by macroscopic stand)  
FS CB: 12 VDC, 18 VA (supplied with external  
power supply unit)

**Tested according to:** UL 61010-1:2012/R:2019-07  
CSA C22.2 No. 61010-1:2012/A1:2018-11

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited Certification body.

**Test report no.:** 713228734

**Date,** 2022-01-26

( Ralph Fischer )