

DE 3 - 31715

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product Laboratory Equipment

(comparison macroscope)

Name and address of the applicant Leica Microsystems CMS GmbH

Ernst-Leitz-Straße 17-37

35578 Wetzlar GERMANY

Name and address of the manufacturer Leica Microsystems CMS GmbH

Ernst-Leitz-Straße 17-37, 35578 Wetzlar, GERMANY

Name and address of the factory

Leica Microsystems CMS GmbH

Ernst-Leitz-Straße 17-37, 35578 Wetzlar, GERMANY

Ratings and principal characteristics Rated voltage: 100-240V Rated frequency: 50/60Hz

Rated frequency: 50/60Hz
Rated power: 120W
Protection class: I

Trade mark (if any) Leica

Model/type Ref. Leica FS M

Additional information (if necessary)

Certificate DE 3 – 3886 issued 2014-08-26 is replaced by this version due to

technical changes

A sample of the product was tested and found

to be in conformity with

IEC 61010-1:2010

IEC 61010-1:2010/AMD1:2016

as shown in the Test Report Ref. No. which forms part of this certificate

713229368

Page 1 of 1

This CB Test Certificate is issued by the National Certification Body

CB 056257 0291 Rev. 00

Date, 2022-02-15

(Erich Thurner)

TÜV SÜD Product Service GmbH • Certification Body • Ridlerstraße 65 • 80339 Munich • Germany



Product Service