




America

CERTIFICATE

No. U8 056257 0290 Rev. 00

Holder of Certificate: **Leica Microsystems CMS GmbH**
Ernst-Leitz-Straße 17-37
35578 Wetzlar
GERMANY

Certification Mark:



Product: **Laboratory equipment
(microscope system)**

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited Certification body.

Test report no.: 713229589

Date, 2022-02-15

(Ralph Fischer)



America

CERTIFICATE

No. U8 056257 0290 Rev. 00

Model(s): **WSU, Supply Unit Scanhead, Pulsed Laser Unit 355 nm, Infinity scanner**

Brand Name: **Leica**

Tested according to: UL 61010-1:2012/R:2019-07
CSA C22.2 No. 61010-1:2012/A1:2018-11
CSA E60825-1:2015

Parameters:

WSU

Rated voltage: 100-240 V
Rated frequency: 50/60 Hz
Rated power: 250 VA
Protection class: I

Supply Unit Scanhead (supplied Infinity Scanner)

Rated voltage: 100-240 V
Rated frequency: 50/60 Hz
Rated power: 150 VA
Protection class: I

Pulsed Laser Unit 355 nm (Ablation laser)

Rated voltage: 12 VDC
(supplied via external limited power supply unit)
Protection class: III
Laser class: 3B
Wavelength: 355 nm
Laser power: 50 mW

Condition of acceptability

- This equipment is for indoor use in non-hazardous locations, operated by qualified personnel skilled in its use.