



America

# CERTIFICATE

No. U8 17 05 56257 162

**Holder of Certificate:** Leica Microsystems CMS GmbH

Ernst-Leitz-Straße 17-37  
35578 Wetzlar  
GERMANY

**Certification Mark:**



**Product:** Laboratory equipment  
(microscope system)

The product was voluntarily tested according to the relevant safety requirements noted above. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited certification body.

**Test report no.:** 028-713099377-000

**Date,** 2017-05-30

Page 1 of 2





America

**CERTIFICATE**  
**No. U8 17 05 56257 162**

**Model(s):** **The microscope system consists of:  
 WSU, Pulsed Laser Unit 355 nm,  
 Infinity scanner, Supply Unit Scanhead**

**Brand Name:** **Leica**

**Parameters:**

WSU	
Rated voltage:	100-240 V
Rated frequency:	50/60 Hz
Rated power:	250 VA
Protection class:	I
Supply Unit Scanhead (supplied Infinity Scanner)	
Rated voltage:	100-240 V
Rated frequency:	50/60 Hz
Rated power:	150 VA
Protection class:	I
Pulsed Laser Unit 355 nm (Ablation laser)	
Rated voltage :	12 VDC (supplied via external limited power supply unit)
Protection class:	III
Laser class:	1
Protection class:	I

**Tested according to:** UL 61010-1:2012/R:2015-07  
 CAN/CSA-C22.2 No. 61010-1:2012 + UPD No. 1:2015-07  
 CAN/CSA-E60825-1:2015-11

**Production Facility(ies):** 56257