



Ref. Certif. No.

DE 3 - 32332

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Laboratory Equipment (Incident Light Stand)
Name and address of the applicant	Leica Microsystems (Schweiz) AG Max Schmidheiny-Str. 201 9435 Heerbrugg SWITZERLAND
Name and address of the manufacturer	Leica Microsystems (Schweiz) AG Max Schmidheiny-Str. 201, 9435 Heerbrugg, SWITZERLAND
Name and address of the factory	Min Aik Technology Co., Ltd 1F,2F,3F,4F,5F,10F,12F,No.492-1,Sec.1., Wan Shou Rd, Wan Shou Rd., Kuei Shan Dist., 33350 Tao Yuan City, TAIWAN
Ratings and principal characteristics	Voltage: 100–240 VAC Frequency: 50/60 Hz Power: max.30 W
Trademark / Brand (if any)	Leica
Model/type Ref.	LED2000 LED2500
A sample of the product was tested and found to be in conformity with	IEC 61010-1:2010 IEC 61010-1:2010/AMD1:2016
as shown in the Test Report Ref. No. which forms part of this certificate	713263207

Page 1 of 1

This CB Test Certificate is issued by the National Certification Body

CB 042849 0013 Rev. 00
Date, 2024-07-26

(Hendrik Elze)

TÜV SÜD Product Service GmbH • Certification Body • Ridlerstraße 65 • 80339 Munich • Germany



Product Service