



Product Service

# CERTIFICATE

No. Z1 056257 0244 Rev. 00

**Holder of Certificate:** **Leica Microsystems CMS GmbH**  
Ernst-Leitz-Straße 17-37  
35578 Wetzlar  
GERMANY

**Certification Mark:**



**Product:** **Laboratory Equipment  
(Microscope)**

**Model(s):**  
Model: DM6 B  
Type: 11889186, 11889187  
Model: DM6 M  
Type: 11889191, 11889192, 11889193, 11889194  
Model: DM6 CS  
Type: 11889199  
Model: Leica LIBS Module  
Type: 11560200

**Parameters:**  
Microscope:  
Equipment is supplied via control unit CTR Compact,  
CTR Compact XY, CTR Advanced or CTR Advanced+.

Leica LIBS Module:  
24 Vdc, 50 W, class III

**Tested according to:** EN 61010-1:2010/A1:2019  
EN 60825-1:2014

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition, the certification holder must not transfer the certificate to third parties. This certificate is valid until the listed date, unless it is cancelled earlier. All applicable requirements of the testing and certification regulations of TÜV SÜD Group have to be complied. For details see: [www.tuvsud.com/ps-cert](http://www.tuvsud.com/ps-cert)

**Test report no.:** 713212608-02  
**Valid until:** 2026-07-22

**Date,** 2021-07-26

( Benedikt Pulver )