



Product Service

CERTIFICATE

No. Z1 18 01 56257 172

Holder of Certificate: **Leica Microsystems CMS GmbH**

Ernst-Leitz-Straße 17-37
35578 Wetzlar
GERMANY

Production Facility(ies):

56257

Certification Mark:



Product:

**Laboratory equipment
(Microscope)**

Model(s):

**DM6 M, DM6 B, DM6 Q, DM6 CS,
Leica LIBS Module**

Parameters:

DM6 B (IVD-version):	11888865, 11888864, 11888863, 11888862
DM6 B:	11888857, 11888859, 11888872, 11888860
DM6 M:	11888777, 11888785, 11888784, 11888776
DM6 Q:	11888880
DM6 CS:	11888413
Leica LIBS Module:	11560200

DM6 B, DM6 M, DM6 Q, DM6 CS: Supplied via control unit
CTR6 LED or control unit CTR6

Leica LIBS Module:

Rated voltage:	24 VDC
Rated power:	50 W
Protection class:	III

Tested according to:

EN 61010-1:2010
EN 61010-2-101:2017
EN 60825-1:2014

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

Test report no.:

028-713112468-001

Valid until:

2023-01-09

Date, 2018-01-10


(Ralph Fischer)

Page 1 of 1

