




America

CERTIFICATE

No. U8 056257 0293 Rev. 00

Holder of Certificate: **Leica Microsystems CMS GmbH**
Ernst-Leitz-Straße 17-37
35578 Wetzlar
GERMANY

Certification Mark:



Product: **Laboratory equipment
(Microscope)**

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL for USA and a Standards Council of Canada ISO/IEC 17065 accredited Certification body for Canada.

Test report no.: 713254136

Date, 2022-03-31

(Ralph Fischer)



America

CERTIFICATE

No. U8 056257 0293 Rev. 00

Model(s):

LMD 6, LMD 7, E-Box LMD 7000

Tested

UL 61010-1:2012/R:2019-07

according to:

UL 61010-2-101:2019

CSA C22.2 No. 61010-1:2012/A1:2018-11

CSA C22.2 No. 61010-2-101:2019

CSA E60825-1:2015

Parameters:**LMD6, LMD7:**

12 VDC, class III

Remark: Supplied via control unit CTR Advanced or control unit CTR Advanced+

E-Box LMD 7000:

100-240 V, 50/60 Hz, 140 VA, class I

Conditions of acceptability:

- This equipment is for indoor use in non-hazardous locations, operated by qualified personnel skilled in its use.
- The microscopes shall be supplied by control unit CTR Advanced or control unit CTR Advanced+.