



America

CERTIFICATE

No. U8 15 09 56257 135

Holder of Certificate: **Leica Microsystems CMS GmbH**

Ernst-Leitz-Straße 17-37
35578 Wetzlar
GERMANY

Production Facility(ies): 56257

Certification Mark:



C US
Laboratory Equipment
(Microscope)

Product:

Model(s):

DMIL M LED
DMIL LED Fluo
(Microscope for In-Vitro-Diagnostic (IVD))
DMIL LED
(Microscope for In-Vitro-Diagnostic (IVD))
DMIL LED Cellfactory
(Microscope for In-Vitro-Diagnostic (IVD))
DMIL LED Fluo Cellfactory
(Microscope for In-Vitro-Diagnostic (IVD))

Parameters:

Rated Voltage: 100-240V
Rated Frequency: 50/60Hz
Rated Input Power: 14 VA
Protection Class: I

Tested according to:

UL 61010-1:2012-05
UL 61010-1:2004/R:2005-07
IEC 61010-2-101:2002
CAN/CSA-C22.2 No. 61010-1:2012-05
CAN/CSA-C22.2 No. 61010-1:2004-07
CAN/CSA-C22.2 No. 61010-2-101:2004-07

The product was voluntarily tested according to the relevant safety requirements noted above. It can be marked with the certification mark above. The mark must not be altered in anyway. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC Guide 67. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited certification body.

Test report no.: 028-713067113-000

Date, 2015-09-08

Page 1 of 1

