



Ref. Certif. No.

DE 3 - 30741

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST
CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)
CB SCHEMESYSTEME CEI D'ACCEPTATION MUTUELLE DE
CERTIFICATS D'ESSAIS DES EQUIPEMENTS
ELECTRIQUES (IECEE) METHODE OC**CB TEST CERTIFICATE
CERTIFICAT D'ESSAI OC**

Product

Laboratory Equipment
(Microscope)

Name and address of the applicant

Leica Microsystems CMS GmbH
Ernst-Leitz-Straße 17-37
35578 Wetzlar, GERMANY

Name and address of the manufacturer

Leica Microsystems Ltd. Shanghai
Building 1, 258 Jinzang Road, China (Shanghai) Pilot Free Trade,
201206 Shanghai, PEOPLE'S REPUBLIC OF CHINA

Name and address of the factory

Leica Microsystems Ltd. Shanghai
Building 1, 258 Jinzang Road, China (Shanghai) Pilot Free Trade,
201206 Shanghai, PEOPLE'S REPUBLIC OF CHINA

Ratings and principal characteristics

Microscope:
Rated voltage: 12V
Rated frequency: DC
Rated input power: 18W
Protection class: III
Lamp classification: Exempt group

Trade mark (if any)

Leica

Model/type Ref.

DM1000 LED

Additional information (if necessary)

Certificate DE 3 – 3630 issued 2013-02-06 is replaced by this
version due to technical changesA sample of the product was tested and found
to be in conformity withIEC 61010-1:2010
IEC 61010-2-101:2015as shown in the Test Report Ref. No.
which forms part of this certificate

028-713135642-000

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**Date, 2018-06-26
CB 18 06 56257 189

Ralph Fischer



TÜV SÜD Product Service GmbH · Certification Body · Ridlerstrasse 65 · D-80339 München

Product Service