



Product Service

# CERTIFICATE

No. Z1 13 02 56257 064

**Holder of Certificate:** Leica Microsystems CMS GmbHErnst-Leitz-Straße 17-37  
35578 Wetzlar  
GERMANY**Production Facility(ies):** 30821**Certification Mark:****Product:** Laboratory Equipment  
(Microscope)**Model(s):** Leica DM1000 LED**Parameters:**

Microscope:	
Rated voltage:	12V
Rated frequency:	DC
Rated input power:	12W
Protection class:	III
LED class:	1

**Tested according to:** EN 61010-1:2001  
EN 61010-2-101:2002  
EN 62471:2008

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

**Test report no.:** 028-71343149-200**Date,** 2013-02-06  
(Ralph Fischer)

Page 1 of 1

