



America

CERTIFICATE


No. U8 13 02 56257 065

Holder of Certificate: Leica Microsystems CMS GmbH

Ernst-Leitz-Straße 17-37
35578 Wetzlar
GERMANY

Production Facility(ies): 30821

Certification Mark:



Product: Laboratory Equipment (Microscope)

Model(s): Leica DM1000 LED

Parameters:

Microscope:	
Rated voltage:	12V
Rated frequency:	DC
Rated input power:	12W
Protection class:	III
LED class:	1

Tested according to:

- UL 61010-1:2004
- IEC 61010-2-101:2002
- CAN/CSA-C22.2 No. 61010-1:2004
- CAN/CSA-C22.2 No. 61010-2-101:2004
- IEC 62471:2006

The product was voluntarily tested according to the relevant safety requirements noted above. It can be marked with the certification mark above. The mark must not be altered in anyway. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC Guide 67. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited certification body.

Test report no.: 028-71343149-200

Date, 2013-02-07

Page i of 1

