



Product Service

CERTIFICATE

No. Z1 15 04 56257 106

Holder of Certificate: Leica Microsystems CMS GmbHErnst-Leitz-Straße 17-37
35578 Wetzlar
GERMANY**Production Facility(ies):**

30821

Certification Mark:**Product:**Laboratory Equipment
(microscope)**Model(s):**DM2500 LED
11888855 (IVD)
11888856 (non-IVD)**Parameters:**

Rated voltage:	100-240 VAC
Rated frequency:	50/60 Hz
Rated power:	40 VA
Protection class:	I

Tested according to:EN 61010-1:2001
EN 61010-1:2010
EN 61010-2-101:2002

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

Test report no.:

028-713058568-000

Valid until:

2020-04-08

Date, 2015-04-13

Page 1 of 1


(Ralph Fischer)
