



America

CERTIFICATE

No. U8 14 08 56257 078

Holder of Certificate: Leica Microsystems CMS GmbH

Ernst-Leitz-Straße 17-37
35578 Wetzlar
GERMANY

Production Facility(ies): 56257

Certification Mark:



Product: Laboratory Equipment (microscope)

Model(s):
DM6000M
DM6000B
(Microscope for In-Vitro-Diagnostic (IVD))
DM5500B
(Microscope for In-Vitro-Diagnostic (IVD))
DM5000B
(Microscope for In-Vitro-Diagnostic (IVD))

Parameters:
Rated voltage: 12 VDC, powered by CTR power supply
Rated power: 100 W
Protection class: III

Tested according to:
UL 61010-1:2004
UL 61010-1:2012
IEC 61010-2-081/A1:2003
IEC 61010-2-101:2002
CAN/CSA-C22.2 No. 61010-1:2004
CAN/CSA-C22.2 No. 61010-2-081:2004
CAN/CSA-C22.2 No. 61010-2-101:2004
CAN/CSA-C22.2 No. 61010-1:2012

The product was voluntarily tested according to the relevant safety requirements noted above. It can be marked with the certification mark above. The mark must not be altered in anyway. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC Guide 67. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited certification body.

Test report no.: 028-713037527-000

Date, 2014-08-07

Page 1 of 1

