



Product Service

CERTIFICATE

No. Z1 14 11 56257 094

Holder of Certificate: Leica Microsystems CMS GmbHErnst-Leitz-Straße 17-37
35578 Wetzlar
GERMANY**Production Facility(ies):** 56257**Certification Mark:****Product:** Laboratory Equipment**Model(s):** FS C (Macroscope stand with table and comparison bridge)
FS CB (comparison bridge)**Parameters:**
FS C: 100-240 V, 50/60 Hz, 125 VA (macroscope stand)
12 VDC, 18 VA (comparison bridge, supplied by microscope stand)
FS CB: 12 VDC, 18 VA (supplied with external power supply unit)**Tested according to:** EN 61010-1:2010

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

Test report no.: 028-713032756-000**Valid until:** 2019-11-09**Date,** 2014-11-12
(Ralph Fischer)

Page 1 of 1

