



Product Service

# C E R T I F I C A T E

No. Z1 14 08 56257 082

**Holder of Certificate:** Leica Microsystems CMS GmbHErnst-Leitz-Straße 17-37  
35578 Wetzlar  
GERMANY**Production Facility(ies):** 56257**Certification Mark:****Product:** Laboratory Equipment  
(macroscope)**Model(s):** Leica FS M**Parameters:**

Rated voltage:	100-240V
Rated frequency:	50/60Hz
Rated power:	120W
Protection class:	I

**Tested according to:** EN 61010-1:2010

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

**Test report no.:** 028-713032744-000**Date,** 2014-08-27

Page 1 of 1

( Ralph Fischer )

