



America

CERTIFICATE

No. Z1 15 07 56257 126

Holder of Certificate: Leica Microsystems CMS GmbH

Ernst-Leitz-Straße 17-37
35578 Wetzlar
GERMANY

Production Facility(ies): 56257

Certification Mark:



Product: Laboratory equipment
(Microscope)

Model(s): LMD 6, LMD 7
LMD 6: 11888871 (IVD), 11888869 (IVD),
11888868, 11888870
LMD 7: 11888876 (IVD), 11888877 (IVD),
11888875, 11888878

Parameters:

Rated voltage:	12 VDC
Protection class:	III
Remark:	Supplied via control unit CTR6 LED or control unit CTR6

Tested according to: EN 61010-1:2010
EN 61010-1:2001
EN 61010-2-101:2002

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

Test report no.: 028-713059304-000

Valid until: 2020-07-29

Date, 2015-07-30 (Ralph Fischer)

Page 1 of 1

