



Product Service

# CERTIFICATE

No. B 14 01 56257 071

**Holder of Certificate:** Leica Microsystems CMS GmbHErnst-Leitz-Straße 17-37  
35578 Wetzlar  
GERMANY**Production Facility(ies):** 56257**Certification Mark:****Product:** Laboratory Equipment  
(Light source)**Model(s):** Leica SFL100**Parameters:**

Rated voltage:	5VDC
Rated power consumption:	10W
Protection class:	III
LED class:	Exempt level

**Tested according to:** EN 61010-1:2010  
EN 62471:2008

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

**Test report no.:** 028-713032753-000**Date,** 2014-01-07  
Page 1 of 1

(Erich Thurner)

