



Ref. Certif. No.

SG PSB-MS-00406

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Laboratory Equipment Microscope
Name and address of the applicant	Leica Microsystems CMS GmbH Ernst-Leitz-Straße 17-37 35578 Wetzlar GERMANY
Name and address of the manufacturer	Leica Microsystems CMS GmbH Ernst-Leitz-Straße 17-37, 35578 Wetzlar, GERMANY
Name and address of the factory	Leica Microsystems Ltd. Shanghai Floor 1, 2, 3A, 4A, and 6, Building T20-1, 258 Jinzang Road, China (Shanghai) Pilot Free Trade Zone, 201206 Shanghai, PEOPLE'S REPUBLIC OF CHINA
Ratings and principal characteristics	Rated voltage: 12V d.c. Rated input: Max. 84W
Model/type Ref.	MATEO FL RUO
Additional information (if necessary)	1. This Microscope should be used with an approved adapter FSP084-DHAN3, the parameters of the adapter: Input: 100-240V~, 1.3A, 50-60Hz; Output: 12V d.c., 7A, 84W; Class I 2. Only the requirements of IEC 61010-1, as documented in the accompanying test report, have been addressed. 3. Include National Differences of United states and Canada
A sample of the product was tested and found to be in conformity with	IEC 61010-1:2010 IEC 61010-1:2010/AMD1:2016
as shown in the Test Report Ref. No. which forms part of this certificate	147-230739-000

Page 1 of 2

This CB Test Certificate is issued by the National Certification Body

CBS 056257 0311 Rev. 00

Date, 2024-04-29

(Dongyun Shen)

TÜV SÜD PSB Pte Ltd 15 International Business Park TÜV SÜD @ IBP Singapore 609937



PSB Singapore



Ref. Certif. No.

SG PSB-MS-00406

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

Trademark / Brand (Image)

Leica

Page 2 of 2

This CB Test Certificate is issued by the National Certification Body

CBS 056257 0311 Rev. 00

Date, 2024-04-29

Dy - sh

(Dongyun Shen)

TÜV SÜD PSB Pte Ltd 15 International Business Park TÜV SÜD @ IBP Singapore 609937



PSB Singapore