

SG PSB-MS-00325

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

**CB TEST CERTIFICATE** 

Product Laboratory Equipment

Microscope

Name and address of the applicant Leica Microsystems

(Suzhou) Technology Co., Ltd. Room 503, Building B2, Genwayl-Park 88 Dongchang Road, Suzhou Industrial Park

215028 Suzhou City, Jiangsu PEOPLE'S REPUBLIC OF CHINA

Name and address of the manufacturer Leica Microsystems (Suzhou) Technology Co., Ltd.

Room 503, Building B2, Genwayl-Park, 88 Dongchang Road, Suzhou Industrial Park, 215028 Suzhou City, Jiangsu, PEOPLE'S REPUBLIC OF

**CHINA** 

Name and address of the factory

Leica Microsystems Ltd. Shanghai

Floor 1, 2, 3A, 4A, and 6, Building T20-1, 258 Jinzang Road, China (Shanghai) Pilot Free Trade Zone, 201206 Shanghai, PEOPLE'S

REPUBLIC OF CHINA

Ratings and principal characteristics Rated voltage: DC 12V

Rated input: MAX.60W For switching power adapter

(model: FSP060-DHAN3, Input:100-240V~, 1.8A, 50-60Hz, Class I; Output:

12.0Vd.c., 5.0A, 60W)

Model/type Ref. MATEO TL RUO

Additional information (if necessary)

Only the requirements of IEC 61010-1, as documented in the accompanying

test report, have been addressed.

National Differences: United states and Canada

A sample of the product was tested and found

to be in conformity with

IEC 61010-1:2010

IEC 61010-1:2010/AMD1:2016

as shown in the Test Report Ref. No. which forms part of this certificate

147-211888-000

Page 1 of 2

This CB Test Certificate is issued by the National Certification Body

CBS 109291 0007 Rev. 00

Date, 2022-08-03

Dy-82

( Dongyun Shen )

TÜV SÜD PSB Pte Ltd 15 International Business Park TÜV SÜD @ IBP Singapore 609937



**PSB** Singapore



SG PSB-MS-00325

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

Trademark / Brand (Image)



Page 2 of 2

This CB Test Certificate is issued by the National Certification Body

CBS 109291 0007 Rev. 00

Date,

2022-08-03

Dy-82



( Dongyun Shen )