



Ref. Certif. No.

DE 3 - 32181

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Laboratory equipment (Ultramicrotome)
Name and address of the applicant	Leica Microsystems CMS GmbH Ernst-Leitz-Straße 17-37 35578 Wetzlar GERMANY
Name and address of the manufacturer	Leica Microsystems CMS GmbH Ernst-Leitz-Straße 17-37, 35578 Wetzlar, GERMANY
Name and address of the factory	Leica Microsystems CMS GmbH Ernst-Leitz-Straße 17-37, 35578 Wetzlar, GERMANY
Ratings and principal characteristics	Rated voltage: 100-240 VAC Rated frequency: 50/60 Hz Rated power: 85 VA Protection class: I
Trademark / Brand (if any)	Leica
Model/type Ref.	UC ENUITY system consist of: UC ENUITY, UC ENUITY Processing unit, UC ENUITY Control unit
A sample of the product was tested and found to be in conformity with	IEC 61010-1:2010 IEC 61010-1:2010/AMD1:2016 IEC 61010-2-081:2019
as shown in the Test Report Ref. No. which forms part of this certificate	713307996-00

Page 1 of 1

This CB Test Certificate is issued by the National Certification Body

CB 056257 0307 Rev. 00

Date, 2023-11-21

Siemon

(Thorsten Siemon)

TÜV SÜD Product Service GmbH • Certification Body • Ridlerstraße 65 • 80339 Munich • Germany



Product Service