


America

CERTIFICATE

No. U8 056257 0308 Rev. 00

Holder of Certificate: Leica Microsystems CMS GmbH
Ernst-Leitz-Straße 17-37
35578 Wetzlar
GERMANY

Certification Mark:



Product: Laboratory equipment
(Ultramicrotome)
Brand Name: Leica

Model(s): UC ENUIITY system consist of:
UC ENUIITY, UC ENUIITY Processing unit,
UC ENUIITY Control unit

Parameters: Rated voltage: 100-240 VAC
Rated frequency: 50/60 Hz
Rated power: 85 VA
Protection class: I

Tested according to: UL 61010-1:2012/R:2019-07
UL 61010-2-081:2019
CSA C22.2 No. 61010-1:2012/A1:2018-11
CSA C22.2 No. 61010-2-081:2019

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL for USA and a Standards Council of Canada ISO/IEC 17065 accredited Certification body for Canada.

Test report no.: 713307996-00

Date, 2023-11-21

Siemon

(Thorsten Siemon)