

1. Manufacturer and or Authorised Representatives Details**Manufacturer**

Name: Leica Microsystems CMS GmbH
Address: Ernst-Leitz-Straße 17-37,
35578 Wetzlar, Germany

2. This declaration is used under the sole responsibility of the manufacturer

We hereby declare that the device described below, both in its basic design and construction and in the version marked by us, conforms to the relevant safety- and health related requirements of the appropriate UK legislations.
This declaration shall cease to be valid if modifications are made to the device without our approval.

3. Object of the declaration

Product: Cryo Chamber
Model: UC Enuity Cryo Chamber
Order no.: 11 708 101

4. First date of compliance

Date: January 16th 2024

5. The products listed are in compliance with the relevant statutory requirements detailed

Electromagnetic Compatibility Regulations 2016

Electrical Equipment (Safety) Regulations 2016

The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012

6. References to the designated standards used to which conformity is declared

BS EN 61010-1+A1:2017	EN 61010-1:2010+A1:2019
BS EN IEC 61010-2-010:2020	EN IEC 61010-2-010:2020
EN IEC 61010-2-011:2022	EN IEC 61010-2-011:2021
BS EN IEC 61326-1:2021	EN 61326-1:2021
BS EN 62471:2009	EN 62471:2008
BS EN IEC 63000:2018	EN IEC 63000:2018

7. Details of Approved Body

n.a.

8. Signed on behalf of the manufacturer by

Name: Steffen Laabs
Job Title: Director RA/QA

Date: 06.12.2023
Signed: 