


America

CERTIFICATE

No. U8 056257 0317 Rev. 00

Holder of Certificate: **Leica Microsystems CMS GmbH**
Ernst-Leitz-Straße 17-37
35578 Wetzlar
GERMANY

Certification Mark:



Product: **Laboratory Equipment
Microscope**

Model(s): **VISORIA B, VISORIA M, VISORIA P**

Parameters:

Rated voltage:	100-240V~
Rated frequency:	50/60Hz
Rated power input:	70VA
Protection class:	I

Tested according to: **UL 61010-1:2012/R:2023-06
CSA C22.2 No. 61010-1:2012/U3:2023-06**

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing, Certification, Validation and Verification Regulations (TCVVR)". TÜV SÜD America Inc. is an OSHA recognized NRTL for USA and a Standards Council of Canada ISO/IEC 17065 accredited Certification body for Canada.

Test report no.: 70.198.24.1902.02-13-0

Date, 2025-03-24

(Shanshan Huang)