



Ref. Certif. No.

DE 3 - 32396

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Laboratory Equipment	
Name and address of the applicant	Leica Microsystems CMS GmbH Ernst-Leitz-Straße 17-37 35578 Wetzlar GERMANY	
Name and address of the manufacturer	Leica Microsystems CMS GmbH Ernst-Leitz-Straße 17-37, 35578 Wetzlar, GERMANY	
Name and address of the factory	Leica Microsystems CMS GmbH Ernst-Leitz-Straße 17-37, 35578 Wetzlar, GERMANY	
Ratings and principal characteristics	Rated Voltage	100 – 240 V ~, 50/60 Hz
	Rated Power	100 VA
	Protection Class	I
Trademark / Brand (if any)	Leica	
Customer's Testing Facility (CTF) Stage used	CTF STAGE 1	
Model/type Ref.	Viventis Deep	
A sample of the product was tested and found to be in conformity with	IEC 61010-1:2010 IEC 61010-1:2010/AMD1:2016 IEC 60825-1:2014	
as shown in the Test Report Ref. No. which forms part of this certificate	713340629	

Page 1 of 1

This CB Test Certificate is issued by the National Certification Body

CB 056257 0314 Rev. 00

Date, 2024-10-25

(Ralph Fischer)

TÜV SÜD Product Service GmbH • Certification Body • Ridlerstraße 65 • 80339 Munich • Germany



Product Service