Ref. Certif. No.

DE 3 - 32796

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Laboratory Equipment (Light sheet microscope system)
Name and address of the applicant	**Leica Microsystems CMS GmbH** Ernst-Leitz-Straße 17-37 35578 Wetzlar GERMANY
Name and address of the manufacturer	Leica Microsystems CMS GmbH Ernst-Leitz-Straße 17-37, 35578 Wetzlar, GERMANY
Name and address of the factory	Leica Microsystems CMS GmbH Ernst-Leitz-Straße 17-37, 35578 Wetzlar, GERMANY
Ratings and principal characteristics	see page 2
Trademark / Brand (if any)	Leica
Customer's Testing Facility (CTF) Stage used	CTF STAGE 1
Model/type Ref.	System: Viventis Scape System consists of following units: 1) SCAPE Module 2) Lightsheet Supply Unit (LSU)
Additional information (if necessary)	Evaluated to National and Group Differences for US, CA.
A sample of the product was tested and found to be in conformity with	IEC 61010-1:2010 IEC 61010-1:2010/AMD1:2016 IEC 60825-1:2014
as shown in the Test Report Ref. No. which forms part of this certificate	713389009

This CB Test Certificate is issued by the National Certification Body

CB 056257 0321 Rev. 00

Date, 2026-03-25

(Martin Heinrich)

TÜV SÜD Product Service GmbH • Certification Body • Ridlerstraße 65 • 80339 Munich • Germany

Product Service

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

Ratings and principal characteristics (continued)

1) SCAPE Module:

Voltage: 24 VDC (SELV supplied by external LSU)
Power: max. 32 W

2) Lightsheet Supply Unit (LSU):

Voltage: 100–240 VAC +/-10%
Frequency: 50 / 60 Hz
Power: max. 175 VA
Protection Class: I
Fuse: T4 A, 250 VAC, breaking capacity H, slow-blowing

This CB Test Certificate is issued by the National Certification Body

CB 056257 0321 Rev. 00
Date, 2026-03-25

(Martin Heinrich)

TÜV SÜD Product Service GmbH • Certification Body • Ridlerstraße 65 • 80339 Munich • Germany

Product Service