ZERTIFIKAT ◆ CERTIFICATE ◆ 認證證書 ◆ СЕРТИФИКАТ ◆ CERTIFICADO ◆ CERTIFICAT

C E R T I F I C A T E

No. U8 056257 0320 Rev. 00

Holder of Certificate: **Leica Microsystems CMS GmbH**
Ernst-Leitz-Straße 17-37
35578 Wetzlar
GERMANY

Certification Mark:

Product: **Laboratory Equipment**
(Light sheet microscope system)

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing, Certification, Validation and Verification Regulations (TCVVR)". TÜV SÜD America Inc. is an OSHA recognized NRTL for USA and a Standards Council of Canada ISO/IEC 17065 accredited Certification body for Canada.

Test report no.: 713389009

Date, 2026-03-25

(Martin Heinrich)

TÜV SÜD America, Inc. • 401 Edgewater Place Suite #500 • Wakefield • MA 01880 • USA

ZERTIFIKAT ◆ CERTIFICATE ◆ 認證證書 ◆ СЕРТИФИКАТ ◆ CERTIFICADO ◆ CERTIFICAT

CERTIFICATE

No. U8 056257 0320 Rev. 00

Model(s): System: Viventis Scape

System consists of following units:
1) SCAPE Module
2) Lightsheet Supply Unit (LSU)

Brand Name: **Leica**

Brand:

Tested according to:
UL 61010-1:2012/R:2024-11
CSA C22.2 No. 61010-1:2012/U4:2024-11
CSA E60825-1:2015

Parameters:

1) SCAPE Module:
Voltage: 24 VDC (SELV supplied by external LSU)
Power: max. 32 W

2) Lightsheet Supply Unit (LSU):
Voltage: 100–240 VAC +/-10%
Frequency: 50 / 60 Hz
Power: max. 175 VA
Protection Class: I
Fuse: T4 A, 250 VAC, breaking capacity H, slow-blowing

TÜV SÜD America, Inc. • 401 Edgewater Place Suite #500 • Wakefield • MA 01880 • USA